

i-ED Coil® Portfolio

A wave of innovations. Infinite treatment options.

kaneka



Infini Coil

Very low memory-shape coil
conforms to irregular spaces

Complex SilkySoft Coil

Our softest finishing coil that fits
through a 0.013" microcatheter

Complex Infini™ Coil

0.012" primary diameter effortlessly seeks space for
volume maximization and to fill with fewer coils

Complex Coil

Our highest memory-shape coil
available in 5 softness profiles



Frame Complex

Order number	Secondary diameter (mm)	Length (cm)	Primary diameter (in)	Softness
392-0304	3	4	0.010	Extra soft
392-0306		6		
392-03H8	3.5	8		
392-0410	4	10	0.012	Soft
392-0510	5	10		
392-0515		15		
392-0620	6	20		
392-0730	7	30		
392-0840	8	40	0.014	Medium soft
392-0940	9	40		
392-1040	10	40		
392-1250	12	50		Standard
392-1450	14	50		
392-1650	16	50		

Filling/Finish Complex-SilkySoft

Order number	Secondary diameter (mm)	Length (cm)	Primary diameter (in)	Softness
390-0101	1	1	0.010	SilkySoft
390-0102		2		
390-0103		3		
390-01H2	1.5	2		
390-01H3		3		
390-01H4		4		
390-0203	2	3		
390-0204		4		
390-0206		6		
390-0208		8		
390-0212		12		
390-02H4	2.5	4		
390-02H6		6		
390-0304	3	4		
390-0306		6		
390-0308		8		

Filling Complex-Infini

Order number	Secondary diameter (mm)	Length (cm)	Primary diameter (in)	Softness
393-020306	2-3	6	0.012	SilkySoft
393-020310		10		
393-030515	3-5	15		
393-030520		20		
393-030530		30		

Multi-utility coil 14-Infini and 10-Infini

Order number	Secondary diameter (mm)	Length (cm)	Primary diameter (in)	Softness
373-040810	4-8	10	0.014	Extra Soft
373-040815		15		
373-040830		30		
373-051230	5-12	30		Soft
373-051250		50		
373-122050	12-20	50		Standard
Order number	Secondary diameter (mm)	Length (cm)	Primary diameter (in)	Softness
371-041010	4-10	10	0.010	Extra Soft
371-041015		15		
372-041020		20		Soft
372-041030		30		



Compatible and indicated for use with 0.013" microcatheters

Kaneka

Kaneka Medical America LLC.

623 Fifth Avenue, 27th Floor,
New York, NY 10022

Email: neuro.vascular@kaneka.com
Web: www.kanekanv.com

© 2025 Kaneka
NVMKT0047
USA_February 2025

RX Only

Caution: US federal law restricts this device to sale by or on the order of a physician. Similar restrictions may be applicable in other countries.

Indications for use: The i-ED Coil® System (i-ED Coil and EDG v4) is intended for the endovascular embolization of intracranial aneurysms and other neurovascular abnormalities such as arteriovenous malformations and arteriovenous fistulae. The i-ED Coil System is also intended for vascular occlusion of blood vessels within the neurovascular system to permanently obstruct blood flow to an aneurysm or other vascular malformation and for arterial and venous embolizations in the peripheral vasculature.

See package insert for complete indications, contraindications, warnings, and instructions for use. For additional information, visit <https://www.kanekanv.com/resources>.

Kaneka® and i-ED Coil® are registered trademarks in the U.S. belonging to Kaneka Corporation. Wave-SR™, Infini™, Complex Infini™, SilkySmooth™, and SilkySoft™ are common law trademarks in the U.S. belonging to Kaneka Corporation.