



Dionex is committed to your success. At the time of installation your Field Service Representative will:

1. Install all instrument gas, liquid, and electric lines required by the chromatograph and pressure test system with required liquids.
2. Dionex will install the application on your system. An application is defined as:
3. Provide instruction on the use of the system with that particular application. This will include operations training and troubleshooting techniques.
4. Instruction on the operation of the Software will be done at this time and suggestions for methods development will be given for future knowledge.
5. Discuss available Technical Assistance from Dionex's Customer Support Organization such as Call Center, Part Number Assistance, Training, and Extended Warranties.

If you have any questions regarding system requirements or need additional information, phone the Call Center at (800) 346-6390.



DX-80 INSTALLATION REQUIREMENTS

- ELECTRICAL:** A standard 120V AC outlet must be located within five feet of the instrument. Laboratories that are subject to frequent power interruptions should give consideration to filtered or uninterruptable power supplies to prevent program loss or possible damage to the equipment.
- GASES:** Reservoirs all require a clean, dry source of helium or nitrogen regulated to 45 psi. All gases must be located within ten (10) feet of the instrument.
- WATER:** A good source of 18.0 M Ω resistivity deionized water must be available for the installation (minimum of 15 liters).
- WASTE:** A waste receptacle must be located within five (5) feet of the instrument. This can be a sink or some other appropriate container.
- CHEMICALS:** Method parameters have been attached which will outline chemical requirements for your installation. It is the customer's responsibility to provide all chemicals and/or standards required for the installation. If these are not available at the time of installation, the installation will have to be rescheduled. A rescheduled installation will be performed on a billable basis and will require a Purchase Order number prior to rescheduling.
- TRAINING:** At installation¹, your Customer Satisfaction Representative will provide operational and preventive maintenance instructions to your primary operator. It is recommended that this person be familiar with the System Operating Manual and PeakNet[®] Tutorial.
- WARRANTY:** Dionex Chromatography Systems are warranted from installation date. However, if you schedule or delay schedule more than sixty (60) days after shipment, the warranty period will begin on the sixty-first (61) day from shipment. See attached Warranty Information Sheet for more details pertaining to your particular system.
- INSTALLATION:** When your system arrives, we ask that you unpack and inspect the equipment for damages and/or any major parts that may be missing from your order. Please notify the shipper of any damages and the Regional Office of any missing parts. Once your system is complete and you have secured all the necessary chemicals for your application, please notify the Call Center at (800) 346-6390.



PeakNet is a registered trademark of Dionex Corporation.

Printed on recycled and recyclable paper.

Dionex Corporation
1228 Titan Way
P.O. Box 3603
Sunnyvale, CA
94088-3603
(408) 737-0700

Dionex Corporation
Salt Lake City Technical Center
1515 West 2200 South, Suite A
Salt Lake City, UT
84119-1484
(801) 972-9292

Dionex U.S. Regional Offices
Sunnyvale, CA (408) 737-8522
Westmont, IL (630) 789-3660
Houston, TX (281) 847-5652
Atlanta, GA (770) 432-8100
Marlton, NJ (856) 596-0600

Dionex International Subsidiaries

Austria (01) 616 51 25 Belgium (32) 3-353 42 93 Canada (905) 844-9650 China (852) 2428 3282 Denmark (45) 36 36 90 90
France (01) 39 30 01 10 Germany (06126) 991-0 Italy (06) 66 51 50 52 Japan (06) 6885-1213 The Netherlands (0161) 43 43 03
Switzerland (062) 205 99 66 United Kingdom (01276) 691722

¹ Designed, developed, and manufactured under an NSAI registered ISO 9001 Quality System.

